

Developing Mass Spectrometry for Water Quantitation and Volatiles Analysis from In-Situ Lunar Regolith

Nilab Azim, Thomas T. Barnes III, Joel A. Olson, Daniel R. Garcia, Abigail A. Eccles, Ray Pitts, Malay Shah, Annie Meier, and Jennifer G. Williams; NASA, Kennedy Space Center, FL (Contact: nilab.azim@nasa.gov)

Introduction: With a revived focus to create a sustainable human presence on the moon, and in preparation for future Mars exploration, it is imperative that all resources are utilized to their fullest potential [1]. In-Situ Resource Utilization (ISRU) will be critical for future mission success as it would enable independent operation of missions while reducing the dependence on the complex supply chain created between the Earth, the moon and Mars.

One of the most critical resources that has been identified for ISRU on the lunar surface is water [2]. Water is a versatile resource that can be used in various space operations, such as radiation shielding and conversion to oxygen (O₂) for propellant [3,4]. One of the most water rich resources is thought to be within the regolith located in the permanently shadowed regions (PSR) of the Moon; several upcoming missions, such as PRIME-1 and VIPER, are already scheduled to confirm this hypothesis [5,6]. However, there are currently no missions planned to quantify the abundance of water at these regions. The quantity of water contained in the ice will inform and help the development of future ISRU plants and operation.

Water and other volatile gases can be detected upon controlled heating of icy lunar regolith for an accurate quantification of water. This method is similar to thermogravimetric analysis, which was used to analyze Apollo lunar regolith samples. However, losses via sublimation/evaporation or contamination may have occurred with the samples due to the transit and exposure to the terrestrial environment. Furthermore, water would not be expected for these samples since they were not collected from a PSR. Thus, for more accurate results, icy lunar regolith should be analyzed *in situ* to provide the most representative composition of both water and other volatile gases.

The Light Water Analysis and Volatile Extract (LightWAVE) project aims to address this knowledge gap. A system has been developed to collect, heat and analyze evolved gases from lunar regolith on the lunar surface. The gases will be analyzed by a modified commercial off the shelf (COTS) residual gas analyzer (RGA) quadrupole mass spectrometer. Evolved gases from heated regolith samples can be identified by the observed mass-to-charge (m/z) ratios, and the peak intensities correlate to the abundances of the observed

gases for quantitation. Therefore, our group is developing a calibration methodology utilizing a COTS RGA to quantitate water from in-situ regolith samples on the lunar surface. The partial pressure of water and ideal gas law will be used to quantify the total water evolved from a regolith sample. Here we present our methodology for the design and calibration for water quantitation using a COTS RGA.

Acknowledgments: We want to thank Aaron Paz, Naina Noorani, Anastasia Ford, and others from NASA JSC for their collaboration on this work. This project is funded by the Space Technology Mission (STMD) Game Changing Development (GCD) Program.

References:

- [1] Artemis Lunar Exploration Program Overview, NASA HQ, <https://www.nasa.gov/specials/artemis/>
- [2] Colaprete, A. et al. (2012) *Space Sci Rev*, 167, 3–22.
- [3] Gibson E. et al. (1974) *Apollo Lunar Sample Analysis*, NASA.
- [4] Mortimer J. et al. (2015) *ICARUS*, 255, pg 3-17.
- [5] <https://www.nasa.gov/viper/overview/>
- [6] https://www.nasa.gov/directorates/spacetech/game_changing_development/projects/PRIME-1